

CONSTRUCTION DIVISION, MATERIALS & PAVEMENTS SECTION (CST-M&P)
AGGREGATE IN-HOUSE CALIBRATION PROCEDURE # 23
Sulfate Soundness Sample Containers

EQUIPMENT CHECKED: Sulfate Soundness Sample Containers (Tex-411-A / ASTM C 88)

1. SCOPE

- 1.1 This procedure provides instructions for verifying the physical condition of sample containers used to determine the SOUNDNESS OF AGGREGATE USING SODIUM SULFATE OR MAGNESIUM SULFATE, Tex-411-A (ASTM C 88).
-

2. APPARATUS

Inspection Equipment Required:

None

Tolerance:

The equipment shall meet the tolerances specified in test procedure Tex-411-A.

3. PROCEDURE

3.1 *Visual Inspection of Sulfate Soundness Sample Containers:*

1. Record the number of the individual sulfate soundness sample container.
2. Visually inspect the container for rust, corrosion, oversized openings, cracks, and/or any other defects that could affect the test results gained by the test (Tex-411-A / ASTM C 88).
3. Repeat for all containers.